

Clear Image Aligners

Models/Scans: Upper and lower models poured in Class I or Class II stone (models). Even if aligners are being requested for a single arch only, it is strongly suggested that complete anterior and posterior anatomy of **both arches** be sent so that any occlusal interferences can be accounted for during fabrication.

Wax Bites: Necessary only if centric occlusion is not obvious when articulating models.

